



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/530,211

04/04/2005

Arnold Keller

246472007600

9918

25227

7590

09/12/2008

MORRISON & FOERSTER LLP

1650 TYSONS BOULEVARD

SUITE 400

MCLEAN, VA 22102

EXAMINER

GANESAN, SUBA

ART UNIT

PAPER NUMBER

3774

MAIL DATE

DELIVERY MODE

09/12/2008

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/530,211

Applicant(s)

KELLER, ARNOLD

Examiner

SUBA GANESAN

Art Unit

3774

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 30 June 2008.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-8 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-8 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-946)
- 3) ☐ Information Disclosure Statement(s) (PTO/SF/ICE)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Response to Arguments

1. Applicant's arguments filed 6/30/2008 have been fully considered but they are not persuasive. Applicant argues that Niederer does not teach or suggest each fin having a surface portion delimited by a medial edge and a lateral edge, and that the medial edge forms a borderline between the surface portion and the steep flank. Examiner disagrees, because an edge is arbitrarily assigned, and thus the fin of Niederer seen from the view of fig. 3 has a medial edge (seen at element 12) that borders the surface portion and the steep flank.
2. Applicant further argues that the fins of Niederer do not project from a face of a shaft. Examiner disagrees, because the shank and neck of the Niederer prosthesis are both considered to comprise a shaft, and thus the location of fin 12 is projecting from a face of a shaft. The width of the fin 12 increases from distal end to the proximal end.
3. Applicant argues that it would not be obvious that the fins extend rectilinearly at an angle of 5-15 degrees with respect to the longitudinal axis of the shaft. Examiner considers this feature to be an obvious design choice that would be considered when designing for ease of insertion into a prepared bone, amount of material used, bone purchase at the implantation site, etc. The level of ordinary skill in the art of hip prosthesis design and use is high, and thus the specific angle of the angled fin of Niederer would have been readily apparent and obvious in view of numerous design considerations for hip prosthesis.

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

2. Claims 1 and 5 are rejected under 35 U.S.C. 102(b) as being anticipated by Niederer (U.S. Pat. No.: 4,359,785).

3. Niederer discloses a hip prosthesis comprising a shaft and a femoral neck, the shaft having a proximal part comprising projecting fins **12** on the front and rear faces of the proximal part (see figs. 1-3). Each fin **12** extends from a distal end of the proximal part to a proximal end of the proximal part (see fig. 2) and has a steep medial flank (see fig. 3), each fin having a surface portion delimited by a medial edge and a lateral edge, the medial edge forming a borderline between the surface portion and the steep flank. The width of the fin increases from the distal end to the proximal end of the proximal part (see fig. 1), and the height of the fins decreases in a lateral direction from an edge delimiting the steep flank (see fig. 3). The prosthesis comprises a device **2** for anchoring the endoprosthesis to a diaphysis.

Claim Rejections - 35 USC § 103

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

1. Claims 4 and 8 are rejected under 35 U.S.C. 103(a) as being unpatentable over Niederer (U.S. Pat. No.: 4,359,785).

2. Niederer is explained supra. Niederer fails to specify whether the height of the lateral edge of the fin is not greater than half the height of the medial edge (See fig. 3). However, it would have been obvious to one of ordinary skill in the art at the time the invention was made to have provided a lateral edge fin height that is not greater than half the height of the medial edge, since it has been held that where the general conditions of a claims are disclosed in the prior art, discovering the optimum or workable ranges involves only routine skill in the art. *In re Aller*, 105 USPQ 233.

3. Claims 2-3 and 6-7 rejected under 35 U.S.C. 103(a) as being unpatentable over Niederer (U.S. Pat. No.: 4,359,785) in view of Tanamal et al (U.S. Pat. No.: 5,755,811).

4. Niederer is explained supra. However, Niederer fails to show the height of the fin above the surface of the shaft increasing from the distal end to the proximal end of the proximal part. Tanamal teaches the use of tapered fins (see fig. 3) resulting in a prosthetic with improved rotational stability (col. 3 lines 32-33). It would have been obvious to one of ordinary skill in the art at the time the invention was made to have provided the finned prosthesis as disclosed by Niederer with a taper towards the distal end of the prosthesis as taught by Tanamal for the purpose of facilitating easier insertion of the finned prosthesis.

5. Tanamal is silent as to the specific angle of the taper. However, it would have been an obvious matter of design choices to have a taper of 5 to 15 degrees, since such a modification would have involved a mere change in size of a component. A change in size is generally recognized as being within the level of ordinary skill in the art. *In re Rose*, 105 USPQ 237 (CCPA 1955). Examiner considers this feature to be an obvious design choice that would be considered when designing for ease of insertion into a prepared bone, amount of material used, bone purchase at the implantation site, etc. The level of ordinary skill in the art of hip prosthesis design and use is high, and thus the specific angle of the angled fin of Niederer would have been readily apparent and obvious in view of numerous design considerations for hip prosthesis.

Conclusion

6. Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to SUBA GANESAN whose telephone number is (571)272-3243. The examiner can normally be reached on M-F 7-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David Isabella can be reached on 571-272-4749. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/S. G./
Examiner, Art Unit 3774
/William H. Matthews/
Primary Examiner, Art Unit 3774